Extraction of a Comprehensive Steroid Panel from Human Serum Using ISOLUTE[®] SLE+ Prior to LC/MS-MS Analysis



Figure 1. Structures of DHEAS, Estradiol and Testosterone.

Introduction

This application note describes the extraction of a panel of steroid hormones from human serum using ISOLUTE" SLE+ Supported Liquid Extraction plates prior to LC/MS-MS analysis. The simple sample preparation procedure delivers clean extracts and high, reproducible recoveries (>75%, RSD <10%) for all analytes in human serum, with linearity >0.99 in the range 5–5000 pg/mL.

ISOLUTE SLE+ Supported Liquid Extraction plates offer an efficient alternative to traditional liquid-liquid extraction (LLE) for bioanalytical sample preparation, providing high analyte recoveries, no emulsion formation, and significantly reduced sample preparation.

Analytes

Cortisol, 18-OH-Corticosterone, 21-Deoxycortisol, Cortisone, Estradiol, 17-OH-Pregnenolone, Aldosterone, 11-Deoxycortisol, Corticosterone, Estrone, DHEA, 17-OH-Progesterone, DHEAS, Testosterone, Dihydrotestosterone (DHT), Pregnenolone, Androstenedione, 11-deoxycorticosterone, Progesterone

Internal standards

DHT-d³ and Aldosterone-d⁴

Sample Preparation Procedure

Format

ISOLUTE[®] SLE+ 400 µL Supported Liquid Extraction plate, part number 820-0400-P01.

Sample Pre-treatment

Add 25 μ L of methanolic ISTD to human serum (total conc. 250 pg/mL). Mix.

Sample Loading

Apply 300 μ L of sample into each well of the ISOLUTE SLE+ plate. Using a Biotage[®] PRESSURE+96 Positive Pressure Manifold, apply 2–5 psi of pressure to load samples onto the sorbent. Wait 5 minutes for the sample to equilibrate on the sorbent.

Analyte Extraction

Apply ethyl acetate (500μ L) and allow to flow under gravity for 5 minutes. Apply a further aliquot of ethyl acetate (500μ L) and allow to flow for 5 minutes under gravity. Apply pressure (5–10 seconds) to remove any remaining extraction solvent.

*this elution procedure should be used when the polar metabolite DHEAS is included in the steroid panel. If DHEAS is not required, and alternative elution solvent system (ethyl acetate: hexane, 75:25, v/v, 2 x 500 µL) should be used. See results section.

Post Elution and Reconstitution

Dry the extract in a stream of air or nitrogen using a Biotage[®] SPE Dry at 40 °C, 20 to 40 L/min, or a TurboVap[®] 96 at 40 °C, 1.5 L/min, for 20 minutes. Reconstitute evaporated samples with 200 μ L of 50:50 (v/v) mobile phase A: mobile phase B, and mix thoroughly.



UHPLC Conditions

Instrument

Shimadzu Nexera UHPLC

Column

ACE C18 (100 mm x 2.1 mm, 1.7 μm)

Mobile Phase

A: 0.2 mM Ammonium Fluoride (aq)

B: Methanol

Flow Rate

Column Temperature

Injection Volume 10 µL

Table 1. UHPLC Gradient

Time (min)	%A	%В
0	50	50
3	40	60
8	10	90
9	5	95
9.1	5	95
9.5	50	50



Mass Spectrometry Conditions

Instrument Shimadzu 8060 Triple Quadrupole MS using ES interface

Nebulizing Gas Flow 3 L/min

Drying Gas Flow 3 L/min

Heating Gas Flow 17 L/min

Interface Temperature 400 °C

DL Temperature 250 °C

Heat Block Temperature 400 °C

CID Gas Flow 270 kPa $\ensuremath{\textbf{Table 2}}.$ MS conditions and retention times for target analytes in positive and negative mode.

Analytes	MRM Transition	Collision Energy	Ion Mode
DHEAS	367.1 > 97.05	33	-
Cortisol	363.4 > 121.25	-24	+
18-OH-Corticosterone	363.3 > 269.2	-16	+
Cortisone	361.3 > 163.15	-22	+
21-Deoxycortisol	347.1 > 311.2	-16	+
Estradiol	271.1 > 145.2	39	-
Aldosterone d ⁴	363.1 > 190.3	19	-
Aldosterone	359.1 > 189.25	18	-
17-OH-Pregnenolone	315.3 > 297.2	-13	+
11-Deoxycortisol	347.3 > 109.25	-27	+
Corticosterone	347.3 > 329.25	-16	+
Estrone	269.2 > 145.2	37	-
11-Deoxycorticosterone	331.3 > 109.05	-25	+
DHEA	289.3 > 253.2	-13	+
Testosterone	289.3 > 97.05	-23	+
DHT-d ³	294.4 > 258.25	-16	+
DHT	291.3 > 255.25	-15	+
Androstenedione	287.3 > 97.2	-21	+
Pregnenolone	299.3 > 159.25	-20	+
17-OH-Progesterone	331.3 > 97.1	-22	+
Progesterone	315.2 > 97.2	-22	+



Results

If the polar metabolite DHEAS is to be included in the steroid suite, ethyl acetate should be used as the elution solvent. If DHEAS is not required, an alternative elution solvent, ethyl acetate: hexane (75:25, v/v) can be used. Recovery data for both elution solvent systems is shown below.

The optimized ISOLUTE® SLE+ protocol using ethyl acetate demonstrated analyte recoveries greater than 75% as shown in Figure 2. The optimized ISOLUTE SLE+ protocol using ethyl acetate:hexane (75:25, v/v) demonstrated analyte recoveries greater than 80% (except DHEAS) as shown in Figure 3. RSDs were below 10% for all analytes using both protocols. Removal of phospholipids was demonstrated by monitoring MRM transitions using the common product ion at m/z 184. Figure 3. Shows the serum phospholipid profile comparing lysophospholipids (bottom trace) and larger molecular weight phospholipids (top trace) extracted using the optimized ISOLUTE SLE+ protocols and 100 μ L of serum protein precipitated with 400 μ L of acetonitrile. Both elution solvents (EtOAc and EtOAc/Hexane(75/25, v/v) eliminated phospholipids compared to a protein precipitated sample.

Recoveries



= 75/25 EtOAc/Hexane EtOAc

Figure 2. Typical analyte % extraction recoveries (n=7) using ethyl acetate (EtOAc) or EtOAc/Hexane (75/25, v/v) as elution solvents.

Protein Precipitation using 100 µL Serum

2. MRM of 11 Chi

els ES-

EtOAc Elution from SLE+

75/25 EtOAc/Hexane Elution from SLE+

2.00 4.00

2.00

4 00

RM of 11 Cha

6.00 8.00

8.00

6.00

Is ES4

4.41e3



6.00



Figure 3. Phospholipid profile comparison between protein precipitated serum and optimized ISOLUTE SLE+ extracts using EtOAc or EtOAc/Hexane (75/25, v/v) as elution solvents.



Figure 4. shows representative chromatography obtained from stripped serum spiked at 5 ng/mL. Satisfactory resolution of the various isobars was obtained using the ACE C18 UPLC column. In order to achieve low level detection of analytes in positive and negative ion modes a combination of 0.2 mM NH4F (aq) and MeOH was utilized.

Linearity was investigated using stripped serum spiked between 5–5000 pg/mL. Good linearity was observed for all analytes typically delivering r² values greater than 0.99. Table 3. details linearity performance and associated LOQ for each analyte using EtOAc and EtOAc/Hexane (75/25, v/v) as elution solvents. Selected calibration curves using EtOAc as elution solvent are shown in Figure 5.



Figure 4. Representative chromatography for analytes spiked at 5 $\mbox{ng/mL}$ in stripped serum.

Table 3. Analyte calibration curve r^2 and LOQ performance.

Analytes	r² EtOAc	LLOQ (pgml EtOAc	.) r ² EtOAc/Hexane 75/25 (v/v)	LLOQ (pg/mL) EtOAc/ Hexane 75/25 (v/v)
DHEAS	0.991	<1000	-	
Cortisol	0.998	< 10	0.998	< 10
18-OH-Corticosterone	0.996	< 100	0.997	< 100
Cortisone	0.992	< 5	0.994	< 5
21-Deoxycortisol	0.997	25	0.998	25
Estradiol	0.996	< 25	0.997	< 25
Aldosterone	0.998	25	0.999	25
17-OH-Pregnenolone	0.999	< 500	0.999	< 500
11-Deoxycortisol	0.999	5	0.998	5
Corticosterone	0.997	< 50	0.999	< 50
Estrone	0.997	10	0.999	10
11-Deoxycorticosterone	0.998	< 10	0.999	< 10
DHEA	0.994	< 1000	0.994	< 1000
Testosterone	0.994	< 5	0.998	< 5
DHT	0.994	< 250	0.999	< 250
Androstenedione	0.990	10	0.998	< 10
Pregnenolone	0.991	< 250	0.998	< 250
17-OH-Progesterone	0.994	< 10	0.998	< 10
Progesterone	0.993	< 5	0.999	< 5





Figure 5. Calibration curves for Estradiol (a), Testosterone (b), 17-OH-Progesterone (c) and Androstenedione (d).

Additional Information

- Ammonium fluoride increased sensitivity in both positive and negative ion modes.
- For increased sensitivity: »
 - » Increase matrix loading volumes to 350 µL
 - Decrease reconstitution solvent volume below 200 µL >>
 - » Increase injection volumes above 10 μ L.
- Steroids can exhibit non-specific binding to plastic **»** collection plates. Different plastics exhibit different binding characteristics. Addition of 2 µL of ethylene glycol to the collection plate prior to evaporation can mitigate this issue. Note: No ethylene glycol was used in generation of the data shown in this application note, utilizing collection plate p/n 121-5203.

Ordering Information

Part Number	Description	Quantity
820-0400-P01	ISOLUTE SLE+ 400 µL Supported Liquid Extraction Plate	1
120-5203	Collection plate, 2 mL square	50
PPM-96	Biotage® PRESSURE+ 96 Positive Pressure Manifold	1
SD-9600-DHS-EU	Biotage® SPE Dry 96 Sample Evaporator 220/240 V	1
SD-9600-DHS-NA	Biotage® SPE Dry 96 Sample Evaporator 100/120 V	1

EUROPE

Main Office: +46 18 565900 Toll Free: +800 18 565710

Order Tel: +46 18 565710

Order Fax: +46 18 565705

Support Tel: +46 18 56 59 11

Fax: +46 18 591922

order@biotage.com

NORTH & LATIN AMERICA

Main Office: +1 704 654 4900 Toll Free: +1 800 446 4752 Fax: +1 704 654 4917 Order Tel: +1 704 654 4900 Order Fax: +1 434 296 8217 ordermailbox@biotage.com Support Tel: +1 800 446 4752 Support Fax: + 46 18 56 57 11 Outside US: +1 704 654 4900 eu-1-pointsupport@biotage.com us-1-pointsupport@biotage.com

IAPAN

Tel: +81 3 5627 3123 Fax: +81 3 5627 3121 jp_order@biotage.com CHINA Tel: +86 21 68162810 Fax: +86 21 68162829 cn_order@biotage.com

KOREA

Tel: + 82 31 706 8500 Fax:+ 82 31 706 8510 korea_info@biotage.com jp-1-pointsupport@biotage.com cn-1-pointsupport@biotage.com kr-1-pointsupport@biotage.com

> Distributors in other regions are listed on www.biotage.com



Part Number: AN890

Part Number: ANS90 © 2018 Biotage. All rights reserved. No material may be reproduced or published without the written permission of Biotage. Information in this document is subject to change without notice and does not represent any commitment from Biotage. E&OE. A list of all trademarks owned by Biotage AB is available at www.biotage.com/legal. Other product and company names mentioned herein may be trademarks or registered trademarks and/or service marks of their respective owners, and are used only for explanation and to the owners' benefit, without intent to infringe. This application note is intended for research use only, not for diagnostic use